



102 E. Bay Front Road
Deale, MD 20751
410-867-6481-Direct
1-877-747-3883-Toll Free

BASIC CALIBRATION DATA SHEET

CALIBRATED BY: E. Heibinger **CUSTOMER:** SAMPLE BASIC DATA SHEET **ENVIRONMENT:**
BALANCE: Mettler XS205 **CONTACT:** _____ **Water:** 21
Serial # B130181039 **PHONE:** 301-761-5012 **INVOICE #:** J0312181A **BP:** 29.82
Due Date 28-Feb-19 **SERVICE DATE:** 12-Mar-18 **DUE DATE:** Mar-19 **Z-Factor:** 1.0031

PIPETTE INFORMATION					REPLACEMENT PARTS			AS CALIBRATED DATA			
PM service performed and all pipettes calibrated in accordance with ISO8655 Specifications.					Shaft	O Ring	Other	LOW	MID	HIGH	PASS / FAIL
Pipette/Serial Number	Manufacturer / Model	Max Volume	As Found (Pass/Fail)	Add'l Svc / Cal Chk							
1	AB61035	GILSON PIPETMAN	2	LOW				0.47	1.02	2.04	PASS
2	B1312925T	RAININ LITE XLS	20	LOW				2.03	10.00	19.92	PASS
3	B1316989T	RAININ LITE XLS	200	LOW		X		20.26	99.43	198.49	PASS
4	BM61374	GILSON PIPETMAN	1000	OUT		X	X	101.03	495.59	993.62	PASS
5	CB59202	GILSON PIPETMAN	100	LOW				10.13	50.16	99.23	PASS
6	H0642471G	RAININ CLASSIC	20	PASS				2.12	10.12	19.84	PASS
7	J0679856A	RAININ LITE	2	LOW				0.46	0.98	1.95	PASS
8	N11135D	GILSON PIPETMAN	1000	PASS				102.06	494.50	993.28	PASS
9	Q58203N	GILSON PIPETMAN	200	PASS				20.54	100.28	201.56	PASS
10	Q64563E	GILSON PIPETMAN	20	LOW				1.97	9.95	20.07	PASS
11	R62024N	GILSON PIPETMAN	200	PASS				20.04	99.64	199.47	PASS
12	S50607H	GILSON PIPETMAN	10	PASS				1.10	5.04	10.01	PASS
13	S66780L	GILSON PIPETMAN	10	LOW				1.06	5.02	10.10	PASS
14	S68095D	GILSON PIPETMAN	200	PASS				20.22	99.32	198.94	PASS
15	S73005G	GILSON PIPETMAN	2	LOW		X		0.52	0.99	1.98	PASS
16	T63416D	GILSON PIPETMAN	10	PASS				1.01	4.99	10.04	PASS
17	U53016L	GILSON PIPETMAN	10	LOW		X		1.10	5.13	9.97	PASS
18	U61127N	GILSON PIPETMAN	200	PASS				20.14	99.73	199.49	PASS
19	U65782H	GILSON PIPETMAN	2	LOW		X		0.45	0.99	2.03	PASS
20	U69922L	GILSON PIPETMAN	1000	PASS				103.62	493.76	993.68	PASS
21	U70071E	GILSON PIPETMAN	20	LOW				2.05	9.93	20.00	PASS
22	U73749L	GILSON PIPETMAN	20	OUT			X	2.06	10.03	20.01	PASS
23	W65323B	GILSON PIPETMAN	1000	PASS				101.41	495.65	996.36	PASS
24	Y52716K	GILSON PIPETMAN	200	PASS				19.92	99.10	198.88	PASS
25	Y57844K	GILSON PIPETMAN	100	PASS				9.95	49.78	99.88	PASS
26	Y58725G	GILSON PIPETMAN	20	HIGH				2.16	10.12	20.12	PASS
27	Z58790D	GILSON PIPETMAN	10	PASS				1.10	4.98	9.92	PASS
28											
29											
30											