

Certificate ID: **F33811**

Certificate of Calibration

Quality Calibrations, Inc.



Device Details

Device ID: 102-444
Serial Number: F33811
Description: Corning Lambda Plus 20
Method: ISO Single Channel
Test Plan: Catalent SC 20
Operator: Judy Nissley

Customer: Sample Customer
Contact: Sample
Location: On Site
Address: Street
City State

119 Lawyers Row
Centreville, MD 21617
Phone: 877-747-3883

Environmental Factors

Air Temperature: 21.20 °C
Barometric Pressure: 30.17 inHg
Relative Humidity: 62 %
Liquid Density: 0.99795
Air Density: 0.00120
Z Factor: 1.00311

Calibration Date: 03-Oct-2023

Status: Passed

Calibration Type: Pipette Calibration

Next Due: 30-Apr-2024

Interval: 6 months

Comment: No Adjustment necessary, 'As found' data serves as 'As left' data.

Measurement Standards

| Device ID | Description | Serial Number | Next Calibration |
|-------------|-------------------|---------------|------------------|
| BALANCE | METTLER XS205DU | B130181039 | 30-Jun-2024 |
| BAROMETER | Digital Barometer | 221461521 | 19-Apr-2024 |
| THERMOMETER | DURAC | J623 | 30-Jun-2026 |

As Found

| Summary Statistics | | | | Accuracy % | | Precision % | | Status | Sample Volumes (µL) | | | | |
|--------------------|-------|------|----------|------------|--------|-------------|--------|--------|---------------------|-------|-------|-------|-------|
| 20.0 uL | Mean | SD | Unc. +/- | Actual | Target | Actual | Target | | 1 | 2 | 3 | 4 | 5 |
| Ch 1 | 19.99 | 0.02 | 0.26 | 0.070 | 1.000 | 0.121 | 0.500 | Passed | 20.01 | 19.97 | 19.95 | 20.00 | 19.99 |
| Ch 1 | 9.98 | 0.03 | 0.26 | 0.231 | 2.000 | 0.262 | 1.000 | Passed | 9.98 | 9.95 | 10.01 | 9.99 | 9.95 |
| Ch 1 | 2.02 | 0.02 | 0.27 | 0.923 | 10.000 | 1.186 | 5.000 | Passed | 2.01 | 2.05 | 2.02 | 1.99 | 2.04 |

Statement of Traceability

This certificate has been issued in accordance with QCI SOP QACCRED and ISO 8655-2 Specifications. This statement certifies that this calibration has been performed with instruments and standards traceable to SI units through NIST or equivalent National Metrology Institute. This certificate shall not be reproduced except in full, without written approval of the laboratory. These results relate only to item calibrated and the results apply to sample as received. QCI uses a simple acceptance approach to the TUR. Customer is to determine if the TUR Accuracy ratio (ratio over uncertainty) is acceptable.



Measurement Uncertainties are determined using the factor of k=2 at a 95% confidence level.

ISO 17025:2017 Certified Certificate AC 1347

Completed by: Judy Nissley

Date: October 3, 2023

Reviewed by: Beverly Heiberger

Date: October 6, 2023